

UV-Vis Spectrophotometer

UV-1900i Plus



User-friendly, Earth-friendly

Since its launch, the UV-1900 series has been widely appreciated not only for its spectral measurement performance, but also for its user-friendly operation and expandability.

The UV-1900i Plus reduces the use of virgin plastics by adopting exterior panels made from resin containing 50 % recycled material, contributing to resource circulation and a reduction in greenhouse gas emissions.

The UV-1900i Plus is an example of Shimadzu Corporation's efforts to provide solutions that support the realization of a decarbonized society.

Connecting Design to the Future

- Realizing a user-friendly and earth-friendly laboratory environment
- Easy-to-use touch-screen display
- Faster response speed with adoption of a new CPU

Various Functions for Comfortable Daily Measurement

- User support with assist and shutdown/wakeup functions

High Performance to Meet Diverse Needs

- The instrument is equipped with an ultra high-speed scan, which can acquire a spectrum in a few seconds, with the lowest level of stray light and noise in its class
- Validation functions enable checks in accordance with Pharmacopoeia (JP/USP/EP) to be performed easily



Connecting Design to the Future

Realizing a user-friendly and earth-friendly laboratory environment

Introduction of sustainable materials

The external panel of the UV-1900i Plus is produced using recycled materials sourced from end-of-life products. This process of resource circulation reduces the use of virgin plastics, contributing to lower greenhouse gas emissions.

Optimization of packaging

By optimizing the volume, weight, and overall structure of product packaging, we help reduce transportation costs and the associated environmental impact.



Shortening system set-up time

By combining the Wakeup function with the Startup validation function the entire process—from system startup to performance verification—can be performed automatically. This automation reduces both user time and labor, as it eliminates the need to remain nearby during setup.


Reduced energy consumption

The Shutdown/Wakeup function enables the system to automatically enter mode after completing an analysis or before starting the next analysis. During sleep mode, energy consumption is minimized, reducing power usage by more than 85 % compared to normal operation (without the Shutdown/Wakeup function).

Easy-to-Use Interface

Grasp the Current Status and Operating Procedures at a Glance

The UV-1900i Plus on-screen user interface includes large, easy-to-see icons deployed on a black background, so the instrument settings are evident at a glance. In addition, the large, easy-to-see icons improve intuitive understanding, which enables users to quickly become familiar with operations. Furthermore, the user interface is designed to minimize transitions between windows, so users do not get confused during operations. The system also features a faster response speed by adopting a new CPU. Display languages are available in eight languages (Japanese, English, Chinese, Spanish, Portuguese, German, French, Russian).



* The background color of the operation screen can also be set to white.



The instrument is equipped with a stylus pen. Operation can be performed with this pen or a finger.

Multiple Measurement Modes

UV-1900i Plus has six basic measurement modes, which offer the optimal solution for a variety of measurements.

Factor L 1.0000				
No.	Sample Name	Abs	K + Abs	8/8
001	SAMPLE1	2.401	2.4000	
002	SAMPLE2	0.101	0.1014	
003	SAMPLE3	0.762	0.7618	
004	SAMPLE4	2.823	2.8234	
005	SAMPLE5	1.213	1.2127	
006	SAMPLE6	1.577	1.5765	
007	SAMPLE7	0.452	0.4517	
008	SAMPLE8	0.203	0.2023	


Photometric

Measures the photometric value at a single wavelength or multiple (up to eight) wavelengths.

Activity Calculation Range: 5.0 s - 40.0 s				
No.	Sample Name	Init (Abs)	Admin Activity	2/2
001	SAMPLE1	0.9386	0.6813	0.5723
002	SAMPLE2	1.0489	0.6451	0.5419


Kinetics

Measures absorbance changes as a function of time, and obtains the enzymatic activity values. The kinetics measurement method or the rate measurement method can be selected.




Spectrum

Measures a sample spectrum using wavelength scanning.



Time Course

Measures changes over time in photometric values at a specified wavelength.



Quantitation

Generates a calibration curve from the measurement of standards, and then calculates the concentrations of unknowns.

Results		3/13
SAMPLE5		
A1(250.0)	= 0.3269	
A2(230.0)	= 0.3094	
Ab(320.0)	= 0.1501	
Abs Ratio = 1.1099		
DNA Conc. = 8.1265		
Protein Conc. = 15.750		

Biomethod


Quantifies DNA or protein concentrations.

Various Functions for Comfortable Daily Measurement

Assist Function



The instrument assists the user to ensure that measurements are performed using the correct procedure.*1





*1 This function can be enabled/disabled.

Shutdown Function

The shutdown function puts UV-1900i Plus to sleep after it's been used, or after a certain period of time. Putting the instrument into sleep mode limits power consumption by the lamps and motors, reducing consumption by more than 85 % compared to the normal standby state.

Wakeup Function

Automatically wakes up a sleeping UV-1900i Plus at a specified time. This function eliminates the need for analysts to wait for it to warm up, allowing them to be more productive.



This product is certified as Shimadzu's Eco-Products Plus. Reduced power consumption by 60 % compared with conventional Shimadzu's products. (Assuming 8 hours operation/day × 5 days/week)




- Automated support functions utilizing digital technology, such as M2M, IoT, and Artificial Intelligence (AI), that enable higher productivity and maximum reliability.
- Allows a system to monitor and diagnose itself, handle any issues during data acquisition without user input, and automatically behave as if it were operated by an expert.
- Supports the acquisition of high quality, reproducible data regardless of an operator's skill level for both routine and demanding applications.


Instrument performance checks can be performed automatically when the instrument is turned ON.*²

It can also be implemented in combination with the Wakeup function.

The performance check items that can be used with this function are as follows: wavelength accuracy (D2), wavelength repeat accuracy (D2), degradation, noise level, baseline flatness, baseline stability (drift), [EP] Control of Wavelength Accuracy (D2), and [USP] Control of Wavelengths (D2).



Data can now be transferred to a PC via a network. With wireless printing, multiple UV units can print from a single printer.*3 (A router and other network equipment must be installed to use a network.)



*5 A PictBridge compatible printer is required.

This saves time when entering sample names for a multi-sample analysis, and prevents sample misidentification and other human errors.*⁶



*6 Use a bar-code reader and keyboard with a USB connection.

Photometric

Ready

USB Wi-Fi

06/26 14:10

Parameters Attachments Photo. Table

K Factor [1.0000]

No.	Sample Name	Abs	K x Abs	2/2
001	SAMPLE1	0.077	0.0773	
002	SAMPLE2	0.093	0.0926	

0.087 Abs

Recalc.

Save Table

Load Table

Print

Auto Zero

Data Comp.

Start


High Performance to Meet Diverse Needs



Low Stray Light

Linearity will be lost in the high absorbance region due to stray light. UV-1900 series is equipped with Low Stray Light Diffraction Gratings, achieve “best-in-class” low stray light. Stray light is at 0.5 % max. (198 nm), making accurate measurements possible at 1 Abs or more, even in the ultraviolet region. In addition, high-concentration samples can be quantified accurately.


The figure on the right is a calibration curve for acetic acid, created with absorbance at 200 nm. The correlation coefficient is 0.9998 and correct measured values are obtained even with high-concentration samples.




Ultra-Fast Scan

Spectra can be acquired as fast as 29,000 nm/min. Ultra-fast scanning is effective for tracking chemical reactions in a short time. In addition to the absorbance change at specified wavelengths, spectra can be acquired in a short time with the UV-1900i Plus. Therefore, more detailed behavior can be investigated by observing spectra with the UV-1900i Plus.

The figures below show the analysis of the particle agglomeration process when salts are added to silver nanoparticles. Measurements of the 300 to 700 nm region were performed in ultra-fast scan mode. In addition to the decrease of absorbance at 400 nm and the increase of absorbance at 480 nm, the temporal changes of spectra can be observed.



Absorption Spectra of Silver Nanoparticles




Temporal Changes of Absorbance at 400 and 480 nm

High Reproducibility and Repeatability Accuracy

The photometric repeatability accuracy is 0.0002 Abs max. (0.5 Abs and 1.0 Abs). As a result, variance in the measurement results is suppressed, enabling more accurate quantitation and the detection of low-concentration samples.

The figure on the right is a calibration curve for caffeine, created with absorbance at 273 nm. The calibration curve has an Abs = 0.0528 Conc. The lower limit of quantitation determined from the standard deviation is 0.0051 mg/L.

No.	Absorbance of Blank Solution (273 nm)
1	-0.00001
2	0.00001
3	-0.00002
4	0.00002
5	0.00001
6	-0.00003
7	0.00001
8	-0.00004
9	0.00001
10	0.00005
Standard Deviation σ	0.000025



Note: One method of determining the lower limit of quantitation is to use ten times the standard deviation. This is an actual measured value and is not guaranteed.

Enables Compliance with ER/ES Regulations and Stronger Data Management


LabSolutions™ UV-Vis Software

Enables higher productivity and provides for a more convenient analytical environment.



Startup

Different measurement applications can be started from the application launcher. Extensive UV validation software programs supporting instrument performance checks as well as optional software programs for special analyses are also available, so various measurements can be intuitively performed.




Four Measurement Mode Windows

Setting Parameters

Instrument Control Panel

Instrument parameter settings can be specified via panels that are separate from the measurement window. The control panels include various functionality that is laid out for superior visibility. Each measurement window connects seamlessly to the corresponding parameter settings window.



Presets (Spectrum and Time Course)

When multiple samples are measured, the sample information can be configured in advance.

Type	Sample Name	Sample No.	# of Peaks	File Name	Parameter	Comments
Sample	SampleName001	Sample001	1	AutoSample001	AutoSample001	
Sample	SampleName002	Sample002	1	AutoSample002	AutoSample002	
Sample	SampleName003	Sample003	1	AutoSample003	AutoSample003	
Sample	SampleName004	Sample004	1	AutoSample004	AutoSample004	
Sample	SampleName005	Sample005	1	AutoSample005	AutoSample005	
Sample	SampleName006	Sample006	1	AutoSample006	AutoSample006	
Sample	SampleName007	Sample007	1	AutoSample007	AutoSample007	
Sample	SampleName008	Sample008	1	AutoSample008	AutoSample008	
Sample	SampleName009	Sample009	1	AutoSample009	AutoSample009	
Sample	SampleName010	Sample010	1	AutoSample010	AutoSample010	
Sample	SampleName011	Sample011	1	AutoSample011	AutoSample011	
Sample	SampleName012	Sample012	1	AutoSample012	AutoSample012	
Sample	SampleName013	Sample013	1	AutoSample013	AutoSample013	
Sample	SampleName014	Sample014	1	AutoSample014	AutoSample014	
Sample	SampleName015	Sample015	1	AutoSample015	AutoSample015	
Sample	SampleName016	Sample016	1	AutoSample016	AutoSample016	
Sample	SampleName017	Sample017	1	AutoSample017	AutoSample017	
Sample	SampleName018	Sample018	1	AutoSample018	AutoSample018	
Sample	SampleName019	Sample019	1	AutoSample019	AutoSample019	
Sample	SampleName020	Sample020	1	AutoSample020	AutoSample020	
Sample	SampleName021	Sample021	1	AutoSample021	AutoSample021	
Sample	SampleName022	Sample022	1	AutoSample022	AutoSample022	
Sample	SampleName023	Sample023	1	AutoSample023	AutoSample023	
Sample	SampleName024	Sample024	1	AutoSample024	AutoSample024	
Sample	SampleName025	Sample025	1	AutoSample025	AutoSample025	
Sample	SampleName026	Sample026	1	AutoSample026	AutoSample026	
Sample	SampleName027	Sample027	1	AutoSample027	AutoSample027	
Sample	SampleName028	Sample028	1	AutoSample028	AutoSample028	
Sample	SampleName029	Sample029	1	AutoSample029	AutoSample029	
Sample	SampleName030	Sample030	1	AutoSample030	AutoSample030	


Instrument Check

Assist Function*



The user can tell at a glance if the instrument has finished warming up. Also, the user is notified when attempting to implement measurements without making necessary corrections, thereby supporting appropriate analysis work.

* This function can be enabled/disabled.



From Measurement to Data Output


Improved Productivity of Data Analysis Operations

Data analysis and data output operations can be performed at the same time as data measurement. Time spent outputting or analyzing data can also be reduced by simultaneously sending data to an Excel® spreadsheet in real time or saving data as text. The software can also automatically perform postprocessing of measured data, such as processing/correcting spectra, and perform pass/fail judgments of measurement results (automatic spectral evaluation).

Automatic Spectral Evaluation (Spectral Evaluation Function)



By specifying various evaluation criteria for measurement results, spectra judgments can be made automatically. In the report creation window, reports can either be prepared based on a previously specified report format or freely laid out based on various parameters, data, or other elements.



Data Management


Robust Data Management Functions

In addition to regular file management in folders on a PC, solutions for saving data in a database with sophisticated security functionality and compliance with ER/ES-related regulations are also available.

Supported Software

LabSolutions DB UV-Vis, LabSolutions CS UV-Vis


Managing data in a database can prevent the overwriting or deletion of analysis data. Furthermore, during postrun analysis, the data can be managed using version numbers, so there are no concerns about overwriting the data.



Shutdown

Shutdown/Wakeup Functions

The UV-1900i Plus shutdown function can also be used from the software. The instrument can shut down or automatically wake up in accordance with settings specified by LabSolutions UV-Vis. After a long period of measurement, the instrument and software can automatically shut down.



Reliable LabSolutions Software

In addition to LabSolutions UV-Vis, which provides basic functionality, Shimadzu offers LabSolutions DB UV-Vis and LabSolutions CS UV-Vis to meet the requirements of ER/ES regulations.


LabSolutions DB UV-Vis

LabSolutions DB UV-Vis allows for secure data management by integrating a data management function with LabSolutions UV-Vis. Compliant with ER/ES regulations, the software is optimally configured for customers using a PC. It is recommended for facilities that do not require network connections and want to be ER/ES compliant.



LabSolutions CS UV-Vis

LabSolutions CS, which is freely accessible to the analysis network, can be connected to LabSolutions UV-Vis, eliminating the need for connecting a PC to the instrument. Since all the data are managed on a server, LabSolutions CS UV-Vis can be read from any personal computer on a network. It is recommended for facilities that have a large number of users, manage data in a database, and want to be ER/ES compliant.




*1 The acquisition controller PC controls analytical instruments.

*2 A terminal server is a server for using terminal services. Users can view data reports and perform electronic signature operations through terminal services. It is ideal for remote connections because of the low network load. Only LC, GC, LCMS, and GCMS support analysis and post-run operations through terminal services.

*3 If a terminal service is used, LabSolutions software does not need to be installed on client PCs or tablets.

*4 Servers can be built on various clouds (IaaS). AWS (Amazon Web Services), Microsoft® Azure®, GCP™ (Google Cloud Platform™)


This instrument can run checks not only for nine JIS items, but also those stipulated in the Japanese Pharmacopoeia (JP), United States Pharmacopoeia (USP), and European Pharmacopoeia (EP). The UV-1900 series also complies with the hardware specifications required by each Pharmacopoeia. In addition, the check conditions can be saved. As a result, checks can be performed easily just by calling them up as needed. Check results can also be saved.



It includes security features that restrict functionality based on user level, allowing three levels of user authority—Administrator, Developer, and Operator—to be assigned to instrument users.

In addition to achieving a resolution of 1 nm, the highest in its class, by using a monochromator with a Czerny-Turner mounting, the UV-1900i Plus also features a compact, bright optical system. The instrument is more than capable of meeting the wavelength resolution required in the USP and EP.

UV validation software (limited function edition) is included as standard. Using this software, inspections can be implemented as stipulated for 9 items in JIS and the JP. If optional UV validation software (USP/EP compatible edition) is added, examinations can be implemented as stipulated in the USP and EP.




- Inspection results can not only be printed but also saved to a file, so the results can be called up later for confirmation.
 - The inspection parameters can also be saved to separate files for periodic and routine inspections, and then called up for use.
 - The user can select confirmation of instrument performance indicators as per JIS K0115 General rules for molecular absorptiometric analysis, as well as the general test methods in the JP. (Order inspection jigs and reagents separately.)
- Note: Optional validation software is required to support the USP and EP.

Applications




Foods

This is an example of the analysis of food dyes. By using ultra-fast scan mode, the time needed for routine spectral checks can be shortened. The 350 nm to 900 nm region can be measured at 1 nm intervals in approx. 4 seconds.





Pharmaceuticals and Life Sciences

This is an example of the analysis of λ DNA. Trace quantities (on the order of a few μ L) can be measured by combining the instrument with Nano Stick and TrayCell[®]. Using TrayCell, a calibration curve for 4 μ L of λ DNA was obtained in the range between 25 ng/ μ L and 500 ng/ μ L.



Chemistry

Materials can be confirmed quantitatively by using LabSolutions UV-Vis and color measurement software. LabSolutions UV-Vis color software shows the chromaticity diagram of the XY color system or Hazen color.



Accessories

Expanded Memory

(P/N 207-23119-41)

The UV-1900i Plus main unit can store up to 999 sets of data. The saved data can be read out from a PC. (A network connection is required for data readout.)

Film Holder

(P/N 204-58909)

Used in transmittance measurement of thin samples such as films and filters. It holds the samples for analysis.



Long-Path Rectangular Cell Holder

(P/N 204-23118-01)

Holds two rectangular cells with an optical path length of 10, 20, 30, 50, 70, or 100 mm.




Test Tube Holder

(P/N 207-23510-41)

Holds test tube instead of sample compartment.

- Outside diameter: $\varnothing 15$ to 22 mm
- Height: 75 to 115 mm

Note: Test tube is not included.



CPS-100 Cell Positioner, Thermoelectrically Temperature Controlled (P/N 206-29500-42, -43, -58)

This attachment permits measurement of up to six sample cells under constant-temperature conditions. Combination of this attachment and the Kinetics mode provides measurement of temperature-sensitive enzyme kinetics of one to six samples.

- Number of cells: 6 on the sample side (temperature-controlled)
1 on the reference side (temperature not controlled)
- Temperature control range: 16 to 60 °C
- Temperature display accuracy (difference from the true value): ± 0.5 °C
- Temperature control precision (variation of temperature): ± 0.1 °C
- Ambient temperature: 15 to 35 °C


Note: Square cells (P/N 200-34442) are not included, please purchase separately.
A USB adapter CPS (P/N 206-25234-91) is required



Multi-Cell Sample Compartment (Type-p)

(P/N 206-69160-58)

Holds up to six 10-mm square cells on the sample side. No temperature control capability.




- Number of cells: 6 on the sample side
1 on the reference side

Note: Square cells are not included, please purchase separately.

Sipper Unit

Model	P/N	Standard Sample Volume
160L (Standard Sipper)	206-23790-51	2.0 mL
160T (Triple-Pass Sipper)	206-23790-52	1.5 mL
160C (Constant-Temperature Sipper)	206-23790-53	2.5 mL
160U (Supermicro Sipper)	206-23790-54	0.5 mL
160V (50 mm Sipper)	206-23790-56	5.0 mL

Four types of sipper units with different flow cells are available. The stepping motordriven peristaltic pump ensures reliable and smooth aspiration of sample solution.



(Direct driving is possible from the UV-1900i Plus so no interface is required.)

Note: The use of a Solenoid Valve (Fluoropolymer) (P/N 204-06599-01) and the SWA-2 Sample Waste Unit (P/N 206-23820-58) are recommended when strong acids, strong alkalis, or organic solvents are to be measured.

TCC-100 Thermoelectrically Temperature Controlled Cell Holder (P/N 206-29510-42, -43, -58)


Uses Peltier effect for controlling the sample and reference temperature, so no thermostated bath or cooling water is required.

- Number of cells: One each on the sample and reference sides (temperature-controlled)
- Temperature control range: 7 to 60 °C
- Temperature display accuracy (difference from the true value): ± 0.5 °C
- Temperature control precision (variation of temperature): ± 0.1 °C

Note: Square cells (P/N 200-34442) are not included, please purchase separately.



Various other accessories are also available. For details, please refer to the UV-Vis Series Accessories (C101-E180).



Analytical Intelligence logo, LabSolutions, PPSQ and SolidSpec are trademarks of Shimadzu Corporation or its affiliated companies in Japan and/or other countries.
Excel, Microsoft and Azure are either registered trademarks or trademarks of Microsoft Corporation in the United States and/or other countries.
Amazon Web Services and AWS are trademarks of Amazon.com, Inc. or its subsidiaries.
Google Cloud Platform and GCP are trademarks of Google LLC.
TrayCell is a registered trademark of Hellma GmbH & Co. KG. in Germany.



Shimadzu Corporation
www.shimadzu.com/an/

For Research Use Only. Not for use in diagnostic procedures.

This publication may contain references to products that are not available in your country. Please contact us to check the availability of these products in your country.

Company names, products/service names and logos used in this publication are trademarks and trade names of Shimadzu Corporation, its subsidiaries or its affiliates, whether or not they are used with trademark symbol "TM" or "®".

Third-party trademarks and trade names may be used in this publication to refer to either the entities or their products/services, whether or not they are used with trademark symbol "TM" or "®".

Shimadzu disclaims any proprietary interest in trademarks and trade names other than its own.

The contents of this publication are provided to you "as is" without warranty of any kind, and are subject to change without notice. Shimadzu does not assume any responsibility or liability for any damage, whether direct or indirect, relating to the use of this publication.